

NADIA ZILLA PARISHAD
KRISHNARAR, NADIA
URL : nadiazillaparishad.gov.in

MEMO NO.- 3153 /NZP

DATE- 16.07.2020

Expression of Interest (EOI)

Expression of Interest is hereby invited from reputed Agencies for purchasing of following medical equipments at Tehatta Sub-divisional Hospital at Tehatta GP under Tehatta –I Block.

Name of equipments with specimen are attached.

Terms & Conditions :

- Bank Draft of 2.5% of total quoted rate as Earnest Money should be submitted along with; EOI in favour of FC & CAO, Nadia Zilla Parishad.
- GST registration certificate & IT returns of recent years will have to be submitted by the agency.
- Credential : 40% of total work value in similar type of work.
- The Machine/Equipments should be covered by free on-site comprehensive Warranty for one year after successful installation
- Any lacuna or lacunae noticed in the functioning of the installation as a result of any design or other features shall be rectified by the supplier free of cost within 72 hours of call log.
- All faults appearing and their rectification shall be periodically advised to the institute, the period being not more than two weeks.
- The bidder will undertake that the supplies of necessary maintenance equipment and spare parts will be made available for all items/equipments and the complete system for at least five years on a continuing basis. However, this does not relive the supplier of any warranty obligations under the contract.

The last date for submitting the expression of Interest (EOI) as per following proforma is 28/07/2020 before 2:00 PM. and the same will be opened on 28 /07 /2020 at 3:00PM.

1. Details of Organization

Name of the Agency	Address	Telephone Number	Contact person's name	Mobile Number	Email ID

2. Total cost for purchase and installation of the Machine/Equipments at Tehatta Sub divisional Hospital at Tehatta.

_____ only inclusive of applicable tax.

3. All supporting documents as mentioned in terms & condition.

The authority reserves the right to accepted / reject rates without assigning any reason thereof.

Sd/-

Secretary
Nadia Zilla Parishad

List of Medical equipments to be purchased under BADP Schemes for Tehatta Sub-divisional Hospital at Tehatta GP under Tehatta –I Block

Sl. No.	Name Of Medical equipments	Specification	Unit Rate	Total
1.	Cardiac Observatory Monitor	<ul style="list-style-type: none"> a) BP, pulse, b) Continuous ECG c) spO2 with curve, d) 24 hr, heart rate trend with defibrillation protection e) Central Monitoring System preferable f) Minimum 1 year warranty. 		
2.	Dental Chair	<ul style="list-style-type: none"> a) Electrically operable Dental Chair with two programme : Zero position & working position b) Ceramic spittoon & tumbler, auto water connection operated by switch c) Halogen light having two intensity: High & Low d) Instrument tray e) Air rotor Control box: One Three way syringe, provision for Air rotor and Micro motor f) Pneumatic Suction g) One Three way syringe for assistant h) Minimum 1 year warranty 		
3.	Pheco Emulsifire Machine (For IOL and Pheco Emulsification)	<ul style="list-style-type: none"> a) Light weight steel hand piece b) Two IA Straight hand pieces units 4 piece. c) Good memory set up for sufficient parameter. d) Multi functional foot switch. e) Input Power 220-240 volt, 50-60 HZ f) Ultrasonic tip frequency, 29-60 KHZ g) Phaco power in both linear and pulse mode. h) Ultrasound pulse vate, 1-14 pulses/sec. i) Micro flow tips. j) Aspiration, 0-500 mmhg linear vaccum k) Multifunctional foot pedal. l) 30 degree US tips: 30 m) Tubing : 6 n) Test Chamsers : 50 o) Diathermy cord: IO p) UPS for minimum one hour back-up q) Minimum 1 year warranty. 		
4	ECG Machine BPL 6208 (12 Lead)	BPL (12 leads)		
5	Trinocular Microscope	<ul style="list-style-type: none"> • Illumination System: 6V 20W halogen lamp (x2) • Observation Tube: Trinocular observation tube (inclination 45), diopter adjustment on the left • Eyepiece (10x): CWHK10X (LB eyepiece 10X), F.N 18mm, anti-fungus (x2) 		
6.	Pharmacy Freezer	180 liter		
7.	Stainless steel medicine almirah	Height 4 feet 5 racks and wheel with net stainless steel		
8.	Blood Cell Separator	Fully automated microprocessor controlled continuous flow cell separator with user friendly touch screen operation. Mobile, easily transportable to patient site for therapeutic uses. It should have high yield leuco-depleted plated collection from a single donor with minimal plasma. It should collect platelet in a presuspended form.		

Copy forwarded for information to :

1. PA to Savadhipati, Nadia Zilla Parishad
2. CA to District Magistrate, Nadia
- 3-6 The S.D.O.....(All).
7. The FC & CAO, Nadia Zilla Parishad .
8. The District Engineer,Nadia Zilla Parishad.
9. The DICO, Nadia . Nadia for wide circulation through 3 nos local news paper
- 10-27 The B.D.O (All).
- 28 DIA , Nadia Zilla Parishad , uploading in website.
- 29 DIO,NIC for uploading in website.
- 30 Notice Board of Nadia Zilla Parishad

Sd/-
Secretary
Nadia Zilla Parishad.